Human IgG-Fc Fragment Cross-Adsorbed Antibody



Goat Polyclonal Conjugate DyLight® 550

Antigen Affinity Purified Catalog No. A80-304D3

Lot No. 15

APPLICATIONS IHC, ICC, Flow Cyt, IF

SPECIES REACTIVITY Human. Minimum reactivity to bovine, chicken, horse, mouse, rabbit, rat and sheep

AMOUNT 1 ml

CONCENTRATION 0.5 mg/ml

STORAGE/SHELF LIFE 2 - 8°C / 1 year from date of receipt

PHYSICAL STATE Liquid

BUFFER Phosphate Buffered Saline (PBS) containing 0.2% BSA and 0.09% Sodium Azide

FLUOROPHORE/PROTEIN 8.5
ISOTYPE IgG
ORIGIN USA

PRODUCTION PROCEDURES

Antiserum was solid phase adsorbed to ensure class specificity. Antiserum was cross adsorbed using bovine, chicken, horse, mouse, rabbit, rat and sheep immunosorbents to remove cross reactive antibodies. The antibody to human IgG was isolated by affinity chromatography using antigen coupled to agarose beads and conjugated to DyLight® 550.

Immunoglobulin concentration was determined using Beer's Law where 1 mg/mL IgG has an A280 of 1.4.

By immunoelectrophoresis and ELISA this antibody reacts specifically with human IgG. Cross reactivity with IgA and IgM is negligible. No antibody was detected against non-immunoglobulin serum proteins. Less than 1% cross reactivity to bovine, chicken, horse, mouse, rabbit, rat and sheep IgG was detected. This antibody may cross react with IgG from

other species.

APPLICATIONS Centrifuge tube to remove product from lid. Optimal working dilutions should be determined

experimentally by the investigator. Prepare working dilution immediately before use.

Immunohistochemistry1:50 - 1:500Immunocytochemistry1:50 - 1:500Flow Cytometry1:50 - 1:200Immunofluorescence1:50 - 1:500

APPLICATION NOTES Not all listed applications have been specifically tested by our laboratory.

DyLight® 550 is excited at 562 (in PBS) and emits at 576 (in PBS). DyLight® 550 replaces DyLight®

549.

DyLight® is a trademark of Thermo Fisher Scientific Inc. and its subsidiaries.

ADDITIONAL INFO https://www.fortislife.com/p/A80-304D3

Use the link above to view SDS, a current list of citations, and other product specific information.

This document certifies that this product has met all of the quality control standards defined by Bethyl Laboratories, Inc.

Michael Spencer, PhD

Date: August 26, 2024

Phone: 800.338.9579 • Fax: 866.597.6105 • Web: www.fortislife.com Orders: orders@fortislife.com • Support: technical@fortislife.com